

Office of the Director
DR. BHUBANESWAR BOROAH CANCER INSTITUTE
A grant-in-aid institute of Department of Atomic Energy, Govt. of India
And a unit of Tata Memorial Centre (Mumbai)
Gopinath Nagar, Guwahati- 781016

No- BCCI-TMC/ 3444 /2023

Dated: 01/08/2023

CORRIGENDUM

The technical specification for the supply, installation & commissioning of **OT Ceiling Mounted Lights 3 arm with camera** at Dr. B. Borooah Cancer Institute against **GeM Bid No: GEM/2023/B/3653984**, dtd. 06/07/2023 has been revised. The revised technical specification for the above mentioned equipment shall be read as follows:

Sl. No	Additional Technical Specification
1	Monitor arm (3rd Arm) of the ceiling mounted light: <ol style="list-style-type: none">1. Monitor arm should have to route the HDMI Cable and power cable from it and it must rotate around 330 degree to its own axis.2. Minimum weight barring capacity of arm should have 20 kg.3. The maximum movement angle of the monitor arm shall be at least up 45 degrees and down 50 degrees.

Other terms & condition of the GeM Bid No: **GEM/2023/B/3653984** shall remain unchanged.

BLA 1/8/23
Chief Administrative Officer
Dr. B. Borooah Cancer Institute
Guwahati – 16

Copy for information to:

1. BCCI website www.bbcionline.org
2. Notice Board, BCCI
3. Concerned file

BLA 1/8/23
Chief Administrative Officer
Dr. B. Borooah Cancer Institute
Guwahati – 16