Office of the Director

DR. B. BOROOAH CANCER INSTITUTE

A Grant- In- Aid Institute of Department of Atomic Energy. Govt. of India And a unit of Tata Memorial Centre, Mumbai Gopinath Nagar, Guwahati-781016 GST No.: 18AAALB0017E1ZW

Date: 13-07-2023

Sub: GeM Tender for Supply, installation and commissioning of Minimal Invasive Set at Dr. B. Borooah Cancer Institute.

Bid No.: GEM/2023/B/3623202, dated: 27.06.2023

CORRIGENDUM

SI No	Existing Tend	ler Item Specifications :	Required Amendments :	Response Against Request
1.	4K UHD Camera	3 times digital zoom	At least 4 times digital zoom lens	No Change in NIT
	Control Uni	lens or more.	or more.	
	(CCU)	4 programmable camera head buttons + device control from camera head. The CCU should have complete accessories so as to connect 4K UHD, FHD, process and	camera hea". The CCU should have complete accessories so as to connect 4K UHD, FHD and 4K in 3D with ICG, process and display the signals in	head buttons + device control from camera head.
2.	4K UHD	display the signals in their native resolution. Request to add /	Should have overlay	
	Camera Control Unit (CCU)	incorporate an additional point under the existing Specifications:	monochromatic and intensity map mode.	No Change in NIT
3.	LED Light Source	Request to include the additional points under	Should be of 400 watts.	No Change in NIT
4.	Forward-	_	Should be compatible with ICG. All telescopes should be ICG	Should be compatible with ICG. All telescopes should be ICG
				enabled.
5.	LED Monitor	Compatible with ICG	Compatible with ICG fluorescent viewing and 3D in 4K.	No Change in NIT

6.	Recorder	Full high definition	41/ digital documentation	TAIZ P. W. L.
0.	Recorder		4K digital documentation system	4K digital documentation system with
		digital documentation	with internal / external media of 2	internal / external media of 2 TB S\$D
		system with external	TB SSD.	
		media of 2 TB SSD.		
7.	High Flow	Size in litres – 45 L	Size in litres – 40 L or more.	Size in litres – 40 L or more
	Insufflator			
8	Hand	Request to add /	Trocar: 5 mm and 10 mm - 02	Trocar: 5 mm and 10 mm - 02 Nos.
	Instruments	incorporate the said	Nos. Each.	Each
		additional Hand	Tooth Occasion 5	Tooth Grasper : 5 mm – 01 No
		Instruments under	Tooth Grasper: 5 mm – 01 No	
		existing Specifications :		
9	Request to add / incorporate include an		All the above equipment's and	All the above equipment's and
	additional point under the existing prescribed Technical Specifications :		instruments should be from one	instruments should be from one
			single manufacturer / same OEM.	single manufacturer / same OEM
10	Conditions :		Physical demonstration of the	Physical demonstration of the quoted
			quoted products in front of the	products in front of the Technical
			Technical Committee is a must for	Committee is a must for all the
			all the bidders.	bidders when required.
11	4K video with 3840 x 2160 pixels resolution.		Camera should be	4K video with 3840 x 2160
			supportable to 3840 x 2160 & 4095 x 2160	pixels resolution or better.
12	Forward oblique Telescope			All the telescopes should be
			UHD (Ultra High	UHD (Ultra High Definition)
			Definition) Telescope.	Telescope.

13419 14/7/23.

Chief Administrative Officer Dr. B. Borooah Cancer Institute Guwahati-16

Сору То,

- 1. Director, BBCI
- 2. GEM Portal
- 3. BBCI Website www.bbcionline.org