

Office of the Director
DR. BHUBANESWAR BOROOAH CANCER INSTITUTE
 A grant-in-aid institute of Department of Atomic Energy, Govt. of India
 And a unit of Tata Memorial Centre (Mumbai)
 Gopinath Nagar, Guwahati- 781016

No- BBCI-TMC/GEM/B/3271/2023

Dated: 14/07/23

CORRIGENDUM

The technical specification for the supply & installation of **Operating microscope for micro vascular surgery (high-end)** at Dr. B. Borooah Cancer Institute against GeM Bid No: **GEM/2023/B/3566459**, dtd. 14/06/2023 has been revised. The revised technical specification for the above mentioned equipment shall be read as follows:

| Sl. No | Present technical specification | Changes made after Pre-bid discussions |
|----------|---|---|
| Point.5 | LIGHT SOURCE: 300W Xenon illumination with integrated 300W Xenon back-up with fast action lamp Changeover. The microscopes' illumination system must provide an additional light beam path to brighten up shadowed areas in the field of view/ Tri LED with 5500K illumination for tumor fluorescence surgery. | No change |
| Point.11 | CAMERA: Fully Integrated 3 CMOS HD Video Camera (Camera not Visible from outside). So that maximum resolution will display & record | CAMERA: Fully Integrated 3 Chip CCD/CMOS HD Video Camera (Camera not Visible from outside). So that maximum resolution will display & record |
| Point.14 | STEREO CO OBSERVER : Should have stereo co observation attachment for side assistant and the attachment should not move in case the head is tilted in forward or backward direction by the main surgeon. The microscope must offer integrated 360° rotatable tube for better ergonomic observation. Binocular should have PD adjustment knob with range of 55mm to 75mm. Binocular should have movement lock in any angle. | Should have stereo co observation attachment for side assistant. The microscope must offer integrated 360° rotatable tube for better ergonomic observation. Binocular should have PD adjustment knob with range of 55mm to 75mm. The binocular should remain lock or fixed unless the microscope is moved of reposition by surgeon. |

| | | |
|---------------------|--|---|
| Point.18 | Sterile microscope covers with one touch drape Air removal /Auto Draping facilities: Draping of the system must be facilitated by an automatic air vacuum/ Auto Drape. 30 no Sterile cover supplied free of cost. | At least 50 no's of sterile microscope cover to be supplied free of cost. |
| Point.19 | DAMPING CORRECTION System should have robotic control active vibration damping mechanism to avoid disturbing vibrations | System should adequate damping correction mechanism to avoid disturbing vibrations. |
| Point.23 | OPTIONAL QUOTE FOR SURGICAL FLUORESCENCE: Surgical Fluorescence: Vascular/Tumor fluorescence mode, where fluorescence objects are emphasized in a ICG and the fluorescence can be observed looking through the eyepiece | SURGICAL FLUORESCENCE: Microscope should have Surgical Fluorescence Vascular & Tumor fluorescence mode. Where fluorescence objects are emphasized in a ICG and the fluorescence can be observed looking through the eyepiece. |
| Point.24 (Addition) | Demo of the quoted model (either at institution or site visit) may be required if so desired by the user after opening of technical bid and prior to the opening of the financial bid. This is for better technical evaluation. | |
| Point.25 (Addition) | Please specify the list of consumables or consumable spares which may need to be replaced Half Yearly or Annually , If any. List of such consumable parts and accessories with their part no and costing must be submitted in the commercial bid. | |
| Point.26 (addition) | Scope of work must include delivery , installation ,demonstration of the complete system and post installation end-user training is must and should be documented which need to be submitted along with installation report. | |
| Point.27 (Addition) | Power cord must be supplied with Indian standard 3 Pin top . | |
| Point.28 (Addition) | The bidder should mentioned the make and model name / number of the quoted equipment and submit the technical brochure of the quoted model in the technical bid along with compliance sheet as per technical specification. | |

The last date of submission of the tender has been extended till **20-07-2023 up to 09:00 AM** and the same will be opened on the same day at **09:30 PM**.

BL 19 14/7/23
Chief Administrative Officer
Dr. B. Borooah Cancer Institute
Guwahati – 16

Copy for information to:

1. BBCI website www.bbcionline.org
2. Notice Board, BBCI
3. Concerned file

BL 19 14/7/23
Chief Administrative Officer
Dr. B. Borooah Cancer Institute
Guwahati – 16