

Office of the Director

DR. B. BOROOAH CANCER INSTITUTE

A Grant- In- Aid Institute of Department of Atomic Energy. Govt. of India
And a unit of Tata Memorial Centre, Mumbai
Gopinath Nagar, Guwahati-781016
GST No.: 18AAALB0017E1ZW

Date: 24-03-2023

Sub: Corrigendum of GeM Tender for Supply, installation & commissioning of 3-chip high definition laparoscopic System at Dr. B. Borooah Cancer Institute.

Bid No.: GEM/2023/B/3195803 dated: 27-02-2023

CORRIGENDUM

Sr. No.	Seller Query/Representation	Corrigendum
1.	HD recorder cum Digital Documentation System for High Definition recording- 1. it should be internal or external HD recorder cum Digital Documentation System for High Definition recording.	No change in NIT. Required internal storage only.
2.	It is a single company specification because already mentioned the company name in the tender specification point no. 8 (The recorder should be compatible to control remotely from existing Stryker 1488 camera head).	By mistakenly a particular brand name is mentioned in the bid. It will be "The recorder should be compatible to control remotely from the camera head".

BL 24/3/23

Sr. Administrative Officer
Dr. B. Borooah Cancer Institute
Guwahati-16