

Dr. Borooh Cancer Institute
A Grant-in-aid institute of Department of Atomic Energy , Govt. Of India
And a unit of Tata Memorial Centre (Mumbai)
Gopinath Nagar, Guwahati-781016
Procurement Department

CORRIGENDUM

Sub: Corrigendum against Micro Motor Drill with following functions Viz. Drill core console, Unidirectional foot switch and burrs. floated against GeM Bid No: GEM/2023/B/3686864

GeM Bid No: GEM/ 2023/B/3686864

SI no.	Technical specification as per NIT	Amendment request	Decision of the committee
A	<u>Motor specs:</u>		
1.	The Drill should have an electric motor driven through a console.	Console should have option multiple hand piece along with the console to have holding capability should have large touch screen monitor	No change in NIT
2.	The stall torque of the motor should be between 4.5-5.5 in -oz	Motor torque: must be in a range of 45mNm ⁻¹ / 3mNm ⁻¹ weight of the drill should not be more than 90gms and length should not be more than 8.5 cm with a diameter not exceeding.	No change in NIT
3.	There should be hand switch option also available	There should be safety switch with console in case of footpedal non function as motor can be hand controlled or foot controlled.	No change in NIT
4.	The chord length should not be less than 4.5meters	The length should not be specific 4.5 meters as it varies from company to company	The chord length should have adequate working length
B	<u>Console specs:</u>		
1	It should be a multispecialty console with provision to attach Neuro Drill, ENT Shaver (micro-debrider), Small bone saws& Heavy duty orthopedic power tools	The system is asked for Neurosurgery, ENT and micro saws as orthopedic require heavy motor would request amendment of orthopedic word	It should be a multispecialty console with provision to attach Neuro Drill, ENT Shaver (micro-debrider), Small bone saws
2	The console should have the provision to use 2 headpieces simultaneously with 2 foot switches	The 2 footswitches is specific to company it should be single pedal footswitch	The console should have the provision to use 2 headpieces simultaneously with 2 foot switches or single footswitch
3	It should be compatible to more than 7 hand pieces	7 hand pieces is very specific to a company should be console should identity all hand piece that is attached	It should be compatible to more than 5 hand pieces

4	The console can be place in any flat surface.	The console should have handle to carry conveniently to any place along this there should be a clamp for holding to iv pole for better placement	This point is omitted from published tender specification
5	There should be curved nose attachment required at least 13CM length/ 5.28 inch length	This is company specific and need to be removed	No change in NIT
6	One Bur should fit all the attachments	There are different burs for different attachments so this is company specific and need to be changed	One Bur should preferably fit all the attachments or different burrs for different attachment

BL/9 30/8/23

Chief Administrative Officer
Dr. B. Borooah Cancer Institute
BB Guwahati-16

Copy to:

1. The Times of India(Mumbai)
(Through APS Advertising)
2. The Assam Tribune
(Through APS Advertising)
3. The Navbharat Times (Delhi)
(Through APS Advertising)
4. Concerned File
5. BBCI Website: www.bbcionline.org

BL/9 30/8/23

Chief Administrative Officer
Dr. B. Borooah Cancer Institute
BB Guwahati-16